ViroGates

K-13-0342-01

# Certificate of Analysis

For Standards and Curve Control used in suPARnostic® AUTO Flex ELISA, code no. E001

**Intended use**

For in vitro diagnostic use. For professional use only.
suPARnostic® standards are intended for establishing standard curve for the quantitative immunological determination of human suPAR (soluble urokinase plasminogen activator receptor) by ELISA.

**Standards a-e**

Recombinant suPAR (soluble urokinase plasminogen activator receptor) in PBS buffer with proprietary additives and 0.05% Bronidox® as preservative. The standards are intended to be used in combination with the reagents provided with suPARnostic AUTO Flex ELISA code E001.

**KIT LOT** ZS 1325 2021-01

| Reagents | | LOT | | Concentration | Range |
|---|---|---|---|---|---|
| Standards | a | ZS 1477 | 2021-01 | 14.5 ng/mL | |
| Standards | b | ZS 1478 | 2021-01 | 11.4 ng/mL | |
| Standards | c | ZS 1479 | 2021-01 | 8.0 ng/mL | |
| Standards | d | ZS 1480 | 2021-01 | 4.4 ng/mL | |
| Standards | e | ZS 1481 | 2021-01 | 0.8 ng/mL | |
| Curve Control | | ZS 1482 | 2021-01 | 2.7 ng/mL | 2.0-3.4 ng/mL |

**PRECAUTIONS AND RECOMMENDATIONS**

Do not use components beyond indicated expiration date.
Do not mix reagents from different kits.

**Calculation of suPAR values from Raw OD´s:** Use the Calculation tool software available on www.virogates.com or contact info@virogates.com

Date of approval: 2019-03-28

Quality Control

Quality Assurance

ViroGates A/S, Blokken 45, DK-3460 Birkerød, Denmark.
Phone: + 45 2113 1336 E-mail: info@virogates.com
Internet: www.virogates.com